



Advancing Transfusion and
Cellular Therapies Worldwide

ASSOCIATION BULLETIN

#06-05

Date: August 30, 2006

UPDATED FEBRUARY 2023

To: AABB Members

From: Christopher D. Hillyer, MD – President
Karen Shoos Lipton, JD – Chief Executive Officer

Re: Monitoring and Preventing the Occurrence of Deviations and Near-Miss
Events in Pretransfusion Testing: Mislabeling/Wrong Blood in Tube

Summary

The ~~33rd~~ ~~24th~~ edition of *Standards for Blood Banks and Transfusion Services*¹ includes language requiring transfusing facilities to monitor deviations and nonconformances (Standard 7.0). Facilities are also required (Standard 8.2) to have a peer review program that monitors, among other issues, patient identification, sample collection and labeling, and near-miss events. This Association Bulletin is intended to provide information to transfusion services on a particularly important and common type of issue associated with pretransfusion testing: the procurement and use of a patient specimen labeled with another individual's name/identification number, referred to as “wrong blood in tube” (WBIT).

Background

A mislabeled blood specimen generally is defined as one whose labeling does not meet the local institutionally defined criteria for accessioning into the laboratory. Common examples include misspelled last names, a missing or incorrect medical record number, or mismatched information between the specimen and the requisition. Such specimens are not suitable for pretransfusion compatibility **testing and the errors associated with them underscore the importance of positive patient identification at the time of sample collection and labeling for safe transfusion.** Guidance on specimen labeling is available in the AABB publication *Guidelines for the Labeling of Specimens for Compatibility Testing*.²

A subset of mislabeling is the problem known as WBIT, where an apparently properly labeled tube identifying blood from one patient actually contains blood from another.³ This type of event is most often recognized when the ABO/Rh result of the current

sample is compared with the historical record on file for the patient. One study that examined prospectively all rejected mislabeled specimens and also noted all discrepant serologic results from “appropriately labeled” samples found that specimens with an obvious labeling error are much more likely to contain WBIT.⁴ **More recent review of FDA fatalities associated with ABO incompatible red cell transfusions show that WBITs, though decreased in frequency, remain one of the most common causes. In all reviewed cases, verification of the ABO type with a second sample or historic type was not performed⁵.**

Incidence of WBIT

Mislabeled specimens, including WBIT, occur frequently. In a study involving one major US academic medical center, Lumadue et al⁴ cite a mislabeling rate of 1 in 71 specimens, and a WBIT rate of 1 in 2800 specimens. Consistent with these observations, a large international prospective study of specimen collection for pretransfusion testing found a mislabeled rate of 1 in 165 specimens and a WBIT rate of approximately 1 in 2000 specimens.^{5,6} **Electronic patient identification systems has been associated with a fivefold reduction in WBIT errors.⁷**

Identified cases of WBIT represent only a subset of the true number of WBIT events because new patients with no historical record are not captured. And, two misidentified patients who share the same blood group by chance are also not captured. Correction factors to account for these two variables may be used to obtain the true WBIT rate from the raw number of WBIT cases identified.³

Conclusion

Identifying the frequency of deviations such as WBIT events will fulfill, in part, the new requirement to monitor blood utilization as specified in Standard 8.2 and at the same time improve the safety of transfusion practices. The new requirement is also consistent with the national patient safety goal of improving the accuracy of patient identification promulgated by the Joint Commission on Accreditation of Healthcare Organizations.

Because transfusion services are already required to check historical blood bank records, any case in which the blood group information does not match the current sample should be identified and investigated. Tracking ~~of~~ these events and root cause analyses will help ~~to~~ develop a corrective action plan to prevent future occurrences. Determining the magnitude of the WBIT specimen problem in an institution is an important step toward the ultimate goal of improving transfusion safety.

Transfusing facilities should take steps to monitor and prevent the occurrence of WBIT. Guidance, including recommendations for monitoring and preventing WBIT, is available in the AABB publication *Guidelines for the Quality Assessment of Transfusion*.^{6,8}

References

1. Gammon R, ed. Standards for blood banks and transfusion services, 33rd edition. Bethesda, MD: AABB, 2022.
2. Rossmann SN for the Scientific Section Coordinating Committee. Guidelines for the labeling of specimens for compatibility testing. Bethesda, MD: AABB, 2002.

3. Dzik WH. Emily Cooley lecture 2002: Transfusion safety in the hospital. *Transfusion* 2003;43:1190-9.

4. Lumadue JA, Boyd JS, Ness PM. Adherence to a strict specimen-labeling policy decreases the incidence of erroneous blood grouping of blood bank specimens. *Transfusion* 1997;37:1169-72.

5. Storch EK, Rogerson B, Eder AF. Trend in ABO-incompatible RBC transfusion-related fatalities reported to the FDA, 2000-2019. *Transfusion* 2020;60:2867-75.

~~5~~6. Dzik WH, Murphy MF, Andreu G, et al. An international study of the performance of specimen collection from patients. *Vox Sang* 2003;85:40-7.

7. Kaufman RM, Dinh AD, et al. Electronic patient identification for sample labeling reduces wrong blood in tube errors. *Transfusion* 2019;59:972-980.

~~6~~8. Wagner J, Saxena S, AuBuchon JP, Shulman IA for the Clinical Transfusion Medicine Committee. Guidelines for the quality assessment of transfusion. Bethesda, MD: AABB, 2006.